
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000364
14:09 20/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:11
STERILIZING 4:15
136.4 C 226kPa 4:15
136.4 C 221kPa 4:45
136.4 C 224kPa 5:15
136.0 C 222kPa 5:45
136.4 C 223kPa 6:15
136.4 C 225kPa 6:45
136.2 C 223kPa 7:15
136.2 C 222kPa 7:45
136.2 C 222kPa 7:45
VENTING 7:45
AIR DRYING 8:36
CYCLE COMPLETE 9:06

Digital Signature #
6C717AC540ED8F28
