

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2972

Name: RITTER 1  
Sterilizer ID: V2178198  
03/06/2023 09:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.1 F 0.0  
2:00 191.2 F 1.7  
4:00 190.5 F 6.0  
6:00 210.8 F 11.5  
8:00 225.4 F 6.5  
10:00 237.6 F 11.8  
12:00 236.2 F 9.2  
14:00 239.5 F 9.3  
16:00 248.6 F 14.4  
18:00 260.8 F 21.3  
20:00 272.0 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.4 F 28.8  
2:00 271.8 F 28.6  
3:00 272.3 F 28.7  
4:00 271.8 F 28.6  
  
Min 271.5 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.7 F 0.0  
5:00 225.3 F 0.0  
10:00 241.9 F 0.0  
15:00 235.8 F 0.0  
20:00 229.6 F 0.0  
25:00 228.6 F 0.0  
30:00 230.1 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 20:31  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:19  
CYCLE COMPLETE

03/06/2023 10:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_