
Midmark M11 - v1.0.0
Total Cycles: 1543

Name: RITTER 1
Sterilizer ID: V2178198
01/20/2022 09:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.9 F 0.0
2:00 176.0 F 2.7
4:00 185.4 F 7.3
6:00 207.4 F 12.5
8:00 219.7 F 6.5
10:00 233.3 F 11.7
12:00 233.9 F 8.6
14:00 240.2 F 10.5
16:00 246.3 F 13.3
18:00 258.7 F 20.0
20:00 269.9 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 28.8
2:00 271.6 F 28.4
3:00 272.5 F 28.8
4:00 271.5 F 28.2

Min 271.5 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 226.0 F 0.0
10:00 236.3 F 0.0
15:00 227.7 F 0.0
20:00 225.1 F 0.0
25:00 220.9 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 20:57
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

01/20/2022 10:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____