
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001968
11:53 16/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:50
PRESSURIZING 1:56
STERILIZING 3:03
136.4 C 225kPa 3:03
136.3 C 224kPa 3:33
136.4 C 224kPa 4:03
136.4 C 223kPa 4:33
136.3 C 226kPa 5:03
136.4 C 226kPa 5:33
136.3 C 224kPa 6:03
136.4 C 226kPa 6:33
136.4 C 226kPa 6:33
VENTING 6:33
AIR DRYING 7:21
CYCLE COMPLETE 21:10

Digital Signature #
03C6AE4AC62C6FCB
