
Midmark M11 - v1.0.5
Total Cycles: 3010

Name: RITTER2
Sterilizer ID: V2313228
10/13/2023 02:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.2 F 0.0
2:00 155.4 F 2.9
4:00 173.0 F 6.7
6:00 197.8 F 11.6
8:00 215.5 F 5.4
10:00 228.0 F 10.4
12:00 231.5 F 8.0
14:00 242.7 F 13.3
16:00 242.7 F 12.3
18:00 255.9 F 19.2
20:00 267.3 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.2
1:00 272.4 F 30.0
2:00 272.0 F 29.2
3:00 272.2 F 29.9
4:00 272.4 F 29.5

Min 271.5 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.2 F 0.0
5:00 219.7 F 0.0
10:00 221.7 F 0.0
15:00 220.3 F 0.0
20:00 218.9 F 0.0
25:00 219.6 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 21:15
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:59:51
CYCLE COMPLETE

10/13/2023 03:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____