

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000981  
10:47 10/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:28  
PRESSURIZING 2:35  
STERILIZING 3:39  
136.4 C 232kPa 3:39  
136.2 C 232kPa 4:09  
136.2 C 231kPa 4:39  
136.0 C 230kPa 5:09  
136.4 C 233kPa 5:39  
136.4 C 233kPa 6:09  
136.4 C 232kPa 6:39  
136.2 C 231kPa 7:09  
136.2 C 231kPa 7:09  
VENTING 7:09  
AIR DRYING 8:00  
CYCLE COMPLETE 11:37  
-----

Digital Signature #  
8977B1314D20B26F  
-----