

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4176

Name: RITTER2  
Sterilizer ID: V2313228  
12/03/2024 11:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.4 F 0.0  
2:00 155.8 F 0.1  
4:00 152.8 F 1.0  
6:00 171.4 F 4.0  
8:00 205.0 F 7.8  
10:00 226.4 F 12.0  
12:00 228.5 F 7.1  
14:00 234.8 F 11.7  
16:00 234.6 F 8.7  
18:00 241.4 F 11.4  
20:00 244.0 F 13.1  
22:00 255.4 F 19.1  
24:00 266.1 F 25.6  
26:00 271.6 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.8  
1:00 272.8 F 30.3  
2:00 271.8 F 29.6  
3:00 272.2 F 30.4  
4:00 272.7 F 30.1

Min 271.5 F 29.6  
Max 273.0 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.0 F 0.0  
5:00 210.6 F 0.0  
10:00 228.8 F 0.0  
15:00 228.3 F 0.0  
20:00 223.9 F 0.0  
25:00 228.6 F 0.0  
30:00 228.9 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 26:38  
STERILIZING: 4:00  
VENTING: 3:58  
DRYING: 30:00  
TOTAL CYCLE: 01:05:33  
CYCLE COMPLETE

12/03/2024 01:00 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_