
Midmark M11 - v1.0.0
Total Cycles: 4142

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2023 03:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.7 F 0.0
2:00 160.7 F 2.8
4:00 177.6 F 6.3
6:00 201.9 F 11.0
8:00 223.2 F 4.8
10:00 232.3 F 9.6
12:00 237.1 F 9.6
14:00 241.2 F 11.5
16:00 239.0 F 9.6
18:00 251.8 F 15.9
20:00 262.9 F 22.2
22:00 272.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.0 F 28.4
2:00 272.5 F 28.6
3:00 272.5 F 28.9
4:00 271.4 F 28.6

Min 271.3 F 28.3
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 232.1 F 0.0
10:00 192.3 F 0.0
15:00 186.7 F 0.0
20:00 182.6 F 0.0
25:00 179.1 F 0.0
30:00 177.6 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 22:55
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:02:06
CYCLE COMPLETE

12/20/2023 04:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____