
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001526
11:00 21/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:48
PRESSURIZING 0:54
STERILIZING 2:02
136.5 C 224kPa 2:02
136.4 C 223kPa 2:32
136.2 C 224kPa 3:02
136.1 C 218kPa 3:32
136.4 C 224kPa 4:02
136.2 C 223kPa 4:32
136.4 C 225kPa 5:02
136.4 C 224kPa 5:32
136.4 C 224kPa 5:32
VENTING 5:32
AIR DRYING 6:23
CYCLE COMPLETE 6:27

Digital Signature #
EEACEAB7F7B4CE90
