

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5121

Name: RITTER 1  
Sterilizer ID: V2178198  
11/21/2024 01:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.3 F 0.0  
2:00 157.4 F 0.2  
4:00 184.4 F 1.1  
6:00 205.7 F 5.6  
8:00 224.5 F 10.0  
10:00 231.4 F 7.8  
12:00 238.4 F 11.7  
14:00 237.0 F 9.7  
16:00 236.6 F 8.7  
18:00 250.1 F 15.4  
20:00 262.3 F 22.3  
22:00 272.5 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.4  
1:00 272.2 F 29.1  
2:00 271.5 F 28.4  
3:00 271.7 F 29.0  
4:00 273.0 F 29.8

Min 271.0 F 28.2  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.7 F 0.0  
5:00 203.6 F 0.0  
10:00 196.4 F 0.0  
15:00 203.2 F 0.0  
20:00 211.1 F 0.0  
25:00 222.3 F 0.0  
30:00 225.5 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 22:18  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:00:48  
CYCLE COMPLETE

11/21/2024 02:55 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_