
Midmark M11 - v1.0.5
Total Cycles: 2392

Name: RITTER2
Sterilizer ID: V2313228
05/18/2023 04:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.6 F 0.0
2:00 183.5 F 4.1
4:00 197.3 F 9.4
6:00 217.9 F 6.5
8:00 231.6 F 11.3
10:00 235.5 F 9.5
12:00 236.5 F 8.8
14:00 250.4 F 15.8
16:00 263.0 F 23.1
18:00 271.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.1 F 29.3
2:00 271.3 F 28.9
3:00 272.2 F 29.4
4:00 271.6 F 28.7

Min 271.2 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 229.2 F 0.0
10:00 225.8 F 0.0
15:00 217.9 F 0.0
20:00 223.6 F 0.0
25:00 229.0 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 18:04
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:57:11
CYCLE COMPLETE

05/18/2023 05:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____