

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1937

Name: RITTER 1  
Sterilizer ID: V2178198  
05/12/2022 08:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 108.2 F 0.0  
2:00 111.1 F 0.0  
4:00 121.6 F 0.8  
6:00 152.8 F 3.0  
8:00 193.9 F 6.8  
10:00 216.6 F 11.4  
12:00 221.8 F 4.5  
14:00 228.4 F 8.9  
16:00 239.4 F 13.1  
18:00 238.4 F 10.4  
20:00 238.5 F 7.9  
22:00 249.1 F 14.8  
24:00 260.9 F 21.4  
26:00 271.7 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.5 F 29.0  
2:00 271.6 F 28.2  
3:00 272.6 F 29.1  
4:00 272.0 F 28.4

Min 271.5 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.3 F 0.0  
5:00 196.9 F 0.0  
10:00 210.6 F 0.0  
15:00 213.1 F 0.0  
20:00 212.2 F 0.0  
25:00 215.0 F 0.0  
30:00 215.8 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 26:35  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:04:32  
CYCLE COMPLETE

05/12/2022 10:01 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02