
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002642
14:37 31/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:13
STERILIZING 3:21
136.4 C 222kPa 3:21
136.4 C 226kPa 3:51
136.4 C 223kPa 4:21
136.3 C 225kPa 4:51
136.3 C 225kPa 5:21
136.0 C 222kPa 5:51
136.4 C 226kPa 6:21
136.3 C 225kPa 6:51
136.3 C 225kPa 6:51
VENTING 6:51
AIR DRYING 7:39
CYCLE COMPLETE 14:36

Digital Signature #
3221C24A2F01484B
