

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2687

Name: RITTER 1  
Sterilizer ID: V2178198  
12/15/2022 04:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.2 F 0.0  
2:00 183.9 F 3.0  
4:00 195.6 F 7.5  
6:00 228.3 F 13.0  
8:00 229.9 F 9.0  
10:00 238.9 F 9.5  
12:00 241.8 F 12.4  
14:00 241.2 F 11.1  
16:00 254.9 F 18.2  
18:00 267.0 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.5 F 29.2  
2:00 271.4 F 28.5  
3:00 272.5 F 29.4  
4:00 272.0 F 28.8  
  
Min 271.4 F 28.5  
Max 272.5 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.8 F 0.0  
5:00 225.0 F 0.0  
10:00 233.3 F 0.0  
15:00 232.9 F 0.0  
20:00 227.5 F 0.0  
25:00 225.5 F 0.0  
30:00 225.8 F 0.0  
  
Drying Complete

FILLING: 1:49  
HEATING: 19:24  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:25  
CYCLE COMPLETE

12/15/2022 05:38 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_