
Midmark M11 - v1.0.0
Total Cycles: 3680

Name: RITTER 1
Sterilizer ID: V2178198
08/21/2023 12:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.0 F 0.0
2:00 180.5 F 1.8
4:00 172.5 F 4.7
6:00 193.4 F 8.6
8:00 207.0 F 13.3
10:00 225.9 F 6.9
12:00 237.0 F 11.4
14:00 233.5 F 7.9
16:00 244.7 F 13.5
18:00 243.5 F 11.8
20:00 255.3 F 17.9
22:00 265.3 F 24.2
24:00 271.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.8 F 29.2
2:00 271.7 F 28.5
3:00 272.4 F 29.3
4:00 271.8 F 28.6

Min 271.2 F 28.4
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.5 F 0.0
5:00 229.5 F 0.0
10:00 228.9 F 0.0
15:00 223.4 F 0.0
20:00 212.4 F 0.0
25:00 225.5 F 0.0
30:00 236.6 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 24:31
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:03:30
CYCLE COMPLETE

08/21/2023 01:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____