

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1348

Name: RITTER 1  
Sterilizer ID: V2178198  
11/09/2021 03:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.7 F 0.0  
2:00 193.7 F 3.8  
4:00 199.8 F 8.7  
6:00 220.7 F 8.4  
8:00 230.1 F 9.6  
10:00 235.2 F 8.8  
12:00 244.4 F 13.4  
14:00 243.9 F 12.4  
16:00 257.3 F 19.5  
18:00 269.1 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.6 F 29.1  
2:00 271.7 F 28.6  
3:00 272.6 F 29.3  
4:00 271.9 F 28.6  
  
Min 271.6 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.7 F 0.0  
5:00 221.2 F 0.0  
10:00 234.2 F 0.0  
15:00 231.6 F 0.0  
20:00 229.4 F 0.0  
25:00 231.0 F 0.0  
30:00 229.3 F 0.0  
  
Drying Complete

FILLING: 2:25  
HEATING: 19:03  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:39  
CYCLE COMPLETE

11/09/2021 04:09 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_