
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001268
7:30 25/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:23
PRESSURIZING 3:25
STERILIZING 4:30
136.5 C 223kPa 4:30
136.4 C 220kPa 5:00
136.1 C 222kPa 5:30
136.2 C 222kPa 6:00
136.3 C 223kPa 6:30
136.0 C 222kPa 7:00
136.4 C 222kPa 7:30
136.4 C 223kPa 8:00
136.4 C 223kPa 8:00
VENTING 8:00
AIR DRYING 8:58
CYCLE COMPLETE 9:05

Digital Signature #
4F56CD261BEEB0C3
