
Midmark M11 - v1.0.5
Total Cycles: 733

Name: RITTER2
Sterilizer ID: V2313228
01/26/2022 02:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.7 F 0.0
2:00 126.3 F 1.6
4:00 137.8 F 4.5
6:00 167.6 F 8.3
8:00 191.9 F 12.9
10:00 217.7 F 6.2
12:00 228.2 F 11.1
14:00 230.5 F 7.2
16:00 241.9 F 13.0
18:00 241.5 F 11.3
20:00 253.6 F 17.6
22:00 264.4 F 24.0
24:00 272.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.2 F 29.6
2:00 272.5 F 29.4
3:00 271.6 F 28.9
4:00 271.1 F 28.9

Min 271.1 F 28.7
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.0 F 0.0
5:00 203.9 F 0.0
10:00 213.8 F 0.0
15:00 221.7 F 0.0
20:00 225.6 F 0.0
25:00 227.0 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 24:11
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:02:34
CYCLE COMPLETE

01/26/2022 03:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____