
Midmark M11 - v1.0.0
Total Cycles: 785

Name: RITTER 1
Sterilizer ID: V2178198
04/15/2021 07:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.2 F 0.0
2:00 74.2 F 0.0
4:00 79.2 F 0.0
6:00 105.2 F 0.0
8:00 149.1 F 0.2
10:00 181.1 F 0.7
12:00 200.6 F 4.1
14:00 213.2 F 7.1
16:00 223.8 F 10.5
18:00 230.2 F 8.2
20:00 234.6 F 10.3
22:00 234.3 F 8.2
24:00 245.7 F 13.2
26:00 243.0 F 11.7
28:00 255.9 F 18.3
30:00 266.9 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.3 F 28.6
2:00 271.8 F 28.6
3:00 271.8 F 28.3
4:00 272.4 F 28.9

Min 271.2 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.4 F 0.0
5:00 231.6 F 0.0
10:00 214.1 F 0.0
15:00 213.6 F 0.0
20:00 210.9 F 0.0
25:00 206.3 F 0.0
30:00 203.8 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:29
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:09:21
CYCLE COMPLETE

04/15/2021 08:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____