

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5197

Name: RITTER 1  
Sterilizer ID: V2178198  
12/17/2024 11:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.5 F 0.0  
2:00 121.9 F 0.0  
4:00 147.3 F 0.0  
6:00 175.9 F 0.1  
8:00 197.6 F 1.1  
10:00 212.5 F 5.3  
12:00 223.9 F 9.3  
14:00 231.7 F 10.6  
16:00 237.3 F 10.2  
18:00 236.2 F 8.4  
20:00 244.1 F 11.7  
22:00 246.6 F 13.3  
24:00 259.2 F 20.1  
26:00 270.6 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.0  
1:00 272.8 F 28.8  
2:00 271.9 F 28.0  
3:00 272.4 F 28.7  
4:00 272.7 F 28.6

Min 271.6 F 27.9  
Max 272.9 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.6 F 0.0  
5:00 201.5 F 0.0  
10:00 207.7 F 0.0  
15:00 211.1 F 0.0  
20:00 210.8 F 0.0  
25:00 214.7 F 0.0  
30:00 211.4 F 0.0  
Drying Complete

FILLING: 0:42  
HEATING: 26:51  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:05:07  
CYCLE COMPLETE

12/17/2024 12:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_