

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5223

Name: RITTER 1  
Sterilizer ID: V2178198  
01/07/2025 01:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 67.9 F 0.0  
2:00 74.9 F 0.0  
4:00 84.4 F 0.0  
6:00 122.6 F 0.0  
8:00 155.8 F 0.2  
10:00 185.6 F 0.3  
12:00 206.6 F 3.6  
14:00 216.8 F 7.0  
16:00 226.8 F 10.5  
18:00 231.0 F 7.9  
20:00 237.7 F 10.4  
22:00 236.0 F 8.0  
24:00 246.3 F 12.7  
26:00 245.1 F 12.6  
28:00 258.1 F 19.5  
30:00 269.6 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.0  
1:00 271.8 F 28.2  
2:00 272.0 F 28.4  
3:00 272.4 F 28.7  
4:00 272.6 F 28.8

Min 271.6 F 27.9  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 229.0 F 0.0  
Dry Cycle Aborted

FILLING: 0:53  
HEATING: 30:51  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 1:19  
TOTAL CYCLE: 00:40:04

01/07/2025 02:20 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_