
Midmark M11 - v1.0.5
Total Cycles: 2843

Name: RITTER2
Sterilizer ID: V2313228
09/07/2023 10:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.1 F 0.0
2:00 167.4 F 2.2
4:00 178.5 F 5.9
6:00 196.1 F 10.2
8:00 220.3 F 6.7
10:00 228.6 F 9.7
12:00 234.8 F 9.5
14:00 240.4 F 12.2
16:00 239.8 F 11.0
18:00 252.0 F 17.1
20:00 262.8 F 23.4
22:00 272.8 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 273.0 F 30.1
2:00 272.8 F 30.0
3:00 272.5 F 30.1
4:00 272.4 F 30.1

Min 270.9 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 209.6 F 0.0
10:00 220.8 F 0.0
15:00 221.9 F 0.0
20:00 220.6 F 0.0
25:00 218.3 F 0.0
30:00 211.1 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 22:23
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:01:25
CYCLE COMPLETE

09/07/2023 12:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____