
Midmark M11 - v1.0.5
Total Cycles: 1675

Name: RITTER2
Sterilizer ID: V2313228
11/03/2022 06:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 80.0 F 0.0
2:00 96.8 F 0.0
4:00 156.4 F 0.0
6:00 184.6 F 0.0
8:00 194.8 F 0.2
10:00 213.2 F 0.4
12:00 221.0 F 3.1
14:00 228.4 F 6.5
16:00 237.6 F 10.2
18:00 238.4 F 10.0
20:00 242.6 F 11.7
22:00 238.8 F 9.7
24:00 238.0 F 9.1
26:00 249.5 F 14.9
28:00 260.8 F 21.3
30:00 271.2 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.7 F 29.3
2:00 272.6 F 29.3
3:00 272.5 F 29.4
4:00 272.1 F 29.3

Min 270.9 F 28.2
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.8 F 0.0
5:00 233.9 F 0.0
10:00 218.2 F 0.0
15:00 226.4 F 0.0
20:00 226.6 F 0.0
25:00 231.6 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 30:33
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:08:37
CYCLE COMPLETE

11/03/2022 07:52 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____