
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003552
8:58 28/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:30
PRESSURIZING 1:33
STERILIZING 2:34
136.5 C 221kPa 2:34
136.4 C 224kPa 3:04
136.2 C 221kPa 3:34
136.1 C 222kPa 4:04
136.4 C 221kPa 4:34
136.3 C 222kPa 5:04
136.2 C 217kPa 5:34
136.4 C 222kPa 6:04
136.4 C 222kPa 6:04
VENTING 6:04
AIR DRYING 7:41
CYCLE COMPLETE 8:00

Digital Signature #
6151CD0278262717
