

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1698

Name: RITTER2  
Sterilizer ID: V2313228  
11/08/2022 12:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.5 F 0.0  
2:00 168.7 F 0.3  
4:00 193.5 F 3.1  
6:00 213.9 F 8.3  
8:00 226.8 F 11.1  
10:00 234.7 F 10.3  
12:00 234.8 F 8.1  
14:00 241.2 F 10.6  
16:00 248.8 F 14.6  
18:00 261.6 F 21.9  
20:00 272.7 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.5 F 29.2  
2:00 271.6 F 28.4  
3:00 271.6 F 29.0  
4:00 272.4 F 29.1

Min 271.0 F 28.2  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.1 F 0.0  
5:00 223.6 F 0.0  
10:00 238.4 F 0.0  
15:00 230.6 F 0.0  
20:00 225.9 F 0.0  
25:00 224.4 F 0.0  
30:00 222.2 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 20:20  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:38  
CYCLE COMPLETE

11/08/2022 01:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_