
Midmark M11 - v1.0.0
Total Cycles: 885

Name: RITTER 1
Sterilizer ID: V2178198
06/10/2021 03:31 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.0 F 0.0
2:00 150.9 F 1.5
4:00 167.3 F 3.9
6:00 188.5 F 7.3
8:00 207.4 F 11.7
10:00 222.7 F 6.0
12:00 233.5 F 10.8
14:00 230.9 F 7.2
16:00 243.9 F 12.9
18:00 242.4 F 11.3
20:00 254.6 F 17.5
22:00 265.2 F 24.0
24:00 272.1 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 271.5 F 28.6
2:00 271.5 F 28.5
3:00 272.0 F 28.7

Min 271.3 F 28.5
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.5 F 0.0
5:00 232.0 F 0.0
10:00 210.9 F 0.0
15:00 211.8 F 0.0
20:00 195.5 F 0.0
25:00 195.9 F 0.0
30:00 188.8 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 24:25
STERILIZING: 3:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:02:05
CYCLE COMPLETE

06/10/2021 04:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____