
Midmark M11 - v1.0.5
Total Cycles: 1531

Name: RITTER2
Sterilizer ID: V2313228
09/26/2022 07:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.7 F 0.0
2:00 77.4 F 0.0
4:00 106.0 F 0.0
6:00 131.9 F 0.0
8:00 158.6 F 1.7
10:00 177.9 F 5.2
12:00 196.3 F 9.1
14:00 211.4 F 13.2
16:00 221.2 F 6.8
18:00 232.8 F 11.1
20:00 233.3 F 8.0
22:00 244.1 F 12.6
24:00 243.8 F 12.6
26:00 256.1 F 19.0
28:00 267.2 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.7 F 29.4
2:00 271.5 F 28.9
3:00 272.5 F 29.3
4:00 271.4 F 28.8

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.1 F 0.0
5:00 190.3 F 0.0
10:00 201.4 F 0.0
15:00 206.1 F 0.0
20:00 196.4 F 0.0
25:00 194.3 F 0.0
30:00 204.5 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 29:26
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:07:22
CYCLE COMPLETE

09/26/2022 08:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____