
Midmark M11 - v1.0.0
Total Cycles: 4359

Name: RITTER 1
Sterilizer ID: V2178198
03/06/2024 09:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.1 F 0.0
2:00 69.6 F 0.2
4:00 76.9 F 1.0
6:00 100.5 F 3.0
8:00 135.8 F 5.8
10:00 162.7 F 9.1
12:00 183.8 F 12.7
14:00 201.8 F 3.3
16:00 212.7 F 6.3
18:00 224.6 F 9.7
20:00 233.5 F 13.5
22:00 233.9 F 9.0
24:00 243.0 F 12.1
26:00 242.5 F 11.5
28:00 254.0 F 17.3
30:00 264.1 F 23.4
32:00 272.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 29.0
2:00 271.5 F 28.2
3:00 272.4 F 29.1
4:00 272.1 F 28.6

Min 271.3 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 208.8 F 0.0
10:00 211.4 F 0.0
15:00 232.3 F 0.0
20:00 232.7 F 0.0
25:00 232.7 F 0.0
30:00 228.5 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:12
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:10:28
CYCLE COMPLETE

03/06/2024 11:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____