
Midmark M11 - v1.0.5
Total Cycles: 3011

Name: RITTER2
Sterilizer ID: V2313228
10/16/2023 08:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.8 F 0.0
2:00 80.4 F 0.0
4:00 77.4 F 0.5
6:00 108.4 F 2.4
8:00 142.6 F 5.3
10:00 170.4 F 8.8
12:00 190.1 F 12.6
14:00 215.1 F 4.1
16:00 223.6 F 7.7
18:00 234.2 F 11.5
20:00 228.8 F 7.1
22:00 242.3 F 12.4
24:00 239.0 F 10.3
26:00 251.3 F 16.3
28:00 261.9 F 22.5
30:00 272.1 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 272.1 F 29.3
2:00 272.7 F 29.5
3:00 272.6 F 29.9
4:00 271.8 F 29.7

Min 271.6 F 29.2
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.9 F 0.0
5:00 233.2 F 0.0
10:00 224.7 F 0.0
15:00 188.1 F 0.0
20:00 183.0 F 0.0
25:00 180.8 F 0.0
30:00 179.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:08
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:09:11
CYCLE COMPLETE

10/16/2023 09:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____