
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005675
7:42 06/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 3:40
STERILIZING 4:45
136.0 C 220kPa 4:45
136.3 C 222kPa 5:15
136.4 C 223kPa 5:45
136.4 C 223kPa 6:15
136.3 C 221kPa 6:45
136.2 C 219kPa 7:15
136.3 C 223kPa 7:45
136.4 C 222kPa 8:15
136.4 C 222kPa 8:15
VENTING 8:15
AIR DRYING 9:05
CYCLE COMPLETE 9:25

Digital Signature #
411A876F66E206AB
