
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006502
9:16 20/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:50
STERILIZING 3:50
136.0 C 220kPa 3:50
136.2 C 223kPa 4:20
136.4 C 220kPa 4:50
136.3 C 221kPa 5:20
136.0 C 222kPa 5:50
136.2 C 221kPa 6:20
136.4 C 223kPa 6:50
136.4 C 223kPa 7:20
136.4 C 223kPa 7:20
VENTING 7:20
AIR DRYING 8:11
CYCLE COMPLETE 21:12

Digital Signature #
0776FAF72448A4F1
