
Midmark M11 - v1.0.5
Total Cycles: 1519

Name: RITTER2
Sterilizer ID: V2313228
09/21/2022 06:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.8 F 0.0
2:00 188.2 F 3.8
4:00 204.8 F 9.2
6:00 218.5 F 8.6
8:00 232.3 F 10.5
10:00 233.8 F 8.1
12:00 241.2 F 11.1
14:00 247.0 F 14.2
16:00 259.8 F 21.2
18:00 271.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.4 F 29.3
2:00 271.9 F 29.2
3:00 272.3 F 29.2
4:00 271.6 F 29.0

Min 271.4 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 208.3 F 0.0
10:00 222.6 F 0.0
15:00 232.9 F 0.0
20:00 227.1 F 0.0
25:00 219.9 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 2:27
HEATING: 18:39
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:16
CYCLE COMPLETE

09/21/2022 07:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____