
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002360
9:23 06/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:14
STERILIZING 3:13
136.0 C 220kPa 3:13
136.0 C 221kPa 3:43
136.4 C 225kPa 4:13
136.4 C 221kPa 4:43
136.3 C 224kPa 5:13
136.4 C 225kPa 5:43
136.3 C 222kPa 6:13
136.4 C 223kPa 6:43
136.4 C 223kPa 6:43
VENTING 6:43
AIR DRYING 7:35
CYCLE COMPLETE 11:33

Digital Signature #
986D6DBA4879431B
