
Midmark M11 - v1.0.5
Total Cycles: 4125

Name: RITTER2
Sterilizer ID: V2313228
11/15/2024 11:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.2 F 0.0
2:00 138.2 F 0.0
4:00 152.1 F 1.8
6:00 181.9 F 5.8
8:00 206.5 F 10.4
10:00 224.3 F 6.2
12:00 234.2 F 10.6
14:00 232.4 F 7.7
16:00 241.3 F 11.5
18:00 245.6 F 13.5
20:00 258.0 F 20.2
22:00 269.0 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.8 F 29.6
2:00 270.9 F 29.0
3:00 273.3 F 30.4
4:00 272.1 F 29.7

Min 270.9 F 28.9
Max 273.7 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.4 F 0.0
5:00 238.7 F 0.0
10:00 228.6 F 0.0
15:00 187.7 F 0.0
20:00 180.2 F 0.0
25:00 176.7 F 0.0
30:00 174.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 23:05
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:01:28
CYCLE COMPLETE

11/15/2024 12:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____