
Midmark M11 - v1.0.5
Total Cycles: 2523

Name: RITTER2
Sterilizer ID: V2313228
06/15/2023 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.3 F 0.0
2:00 166.8 F 3.5
4:00 179.0 F 7.5
6:00 198.8 F 12.5
8:00 219.4 F 7.3
10:00 232.7 F 12.7
12:00 235.0 F 10.0
14:00 236.4 F 8.8
16:00 248.0 F 14.9
18:00 259.7 F 21.5
20:00 270.3 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.3
1:00 272.3 F 30.1
2:00 271.9 F 29.4
3:00 271.2 F 29.5
4:00 272.9 F 30.1

Min 271.1 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 221.3 F 0.0
10:00 230.6 F 0.0
15:00 228.4 F 0.0
20:00 229.9 F 0.0
25:00 225.7 F 0.0
30:00 222.4 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 20:57
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:59:40
CYCLE COMPLETE

06/15/2023 02:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____