
Midmark M11 - v1.0.0
Total Cycles: 2803

Name: RITTER 1
Sterilizer ID: V2178198
01/23/2023 10:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.2 F 0.0
2:00 167.5 F 0.0
4:00 171.6 F 2.9
6:00 204.6 F 7.4
8:00 217.8 F 12.8
10:00 225.1 F 7.4
12:00 237.1 F 12.6
14:00 237.2 F 9.9
16:00 238.5 F 8.5
18:00 248.6 F 14.6
20:00 260.6 F 21.4
22:00 271.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.2 F 29.3
2:00 272.2 F 28.9
3:00 271.4 F 28.4
4:00 271.6 F 28.8

Min 271.1 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 213.9 F 0.0
10:00 231.3 F 0.0
15:00 232.4 F 0.0
20:00 227.6 F 0.0
25:00 225.6 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:41
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:58
CYCLE COMPLETE

01/23/2023 11:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____