

-----  
Midmark M11 - v1.0.0  
Total Cycles: 995

Name: RITTER 1  
Sterilizer ID: V2178198  
07/12/2021 11:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.0 F 0.0  
2:00 171.5 F 2.3  
4:00 178.3 F 5.5  
6:00 208.4 F 10.1  
8:00 221.4 F 5.5  
10:00 229.9 F 8.9  
12:00 242.0 F 12.0  
14:00 239.9 F 10.9  
16:00 237.5 F 8.9  
18:00 250.7 F 15.5  
20:00 261.9 F 21.8  
22:00 272.1 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.5 F 28.8  
2:00 271.5 F 28.0  
3:00 272.6 F 28.9  
4:00 272.1 F 28.4

Min 271.4 F 28.0  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.3 F 0.0  
5:00 231.2 F 0.0  
10:00 226.7 F 0.0  
15:00 227.7 F 0.0  
20:00 225.5 F 0.0  
25:00 230.1 F 0.0  
30:00 229.1 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 22:32  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:46  
CYCLE COMPLETE

07/12/2021 12:10 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01