

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1534

Name: RITTER 1  
Sterilizer ID: V2178198  
01/17/2022 03:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.7 F 0.0  
2:00 121.9 F 0.6  
4:00 159.1 F 3.5  
6:00 184.8 F 8.1  
8:00 202.0 F 13.4  
10:00 216.2 F 6.6  
12:00 227.7 F 11.3  
14:00 230.2 F 7.7  
16:00 243.5 F 13.2  
18:00 242.2 F 12.0  
20:00 255.3 F 18.6  
22:00 266.8 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.1 F 29.2  
2:00 272.0 F 29.4  
3:00 271.8 F 29.0  
4:00 272.3 F 29.6

Min 271.3 F 28.7  
Max 272.5 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.8 F 0.0  
5:00 216.6 F 0.0  
10:00 232.1 F 0.0  
15:00 232.3 F 0.0  
20:00 229.1 F 0.0  
25:00 228.5 F 0.0  
30:00 229.0 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 23:26  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:01:32  
CYCLE COMPLETE

01/17/2022 04:47 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_