
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003979
8:16 11/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:10
STERILIZING 4:14
136.5 C 221kPa 4:14
136.1 C 221kPa 4:44
136.2 C 221kPa 5:14
136.4 C 222kPa 5:44
136.4 C 219kPa 6:14
136.4 C 221kPa 6:44
136.4 C 220kPa 7:14
136.4 C 221kPa 7:44
136.4 C 221kPa 7:44
VENTING 7:44
AIR DRYING 8:33
CYCLE COMPLETE 8:45

Digital Signature #
9A5F67CF1D264531
