
Midmark M11 - v1.0.5
Total Cycles: 4423

Name: RITTER2
Sterilizer ID: V2313228
03/13/2025 09:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.6 F 0.0
2:00 76.9 F 0.0
4:00 71.9 F 0.3
6:00 80.7 F 2.0
8:00 133.0 F 5.0
10:00 161.0 F 7.6
12:00 181.8 F 10.5
14:00 196.0 F 8.8
16:00 209.2 F 5.7
18:00 219.2 F 9.2
20:00 229.6 F 13.1
22:00 232.7 F 9.0
24:00 242.5 F 12.2
26:00 242.0 F 11.9
28:00 252.9 F 17.4
30:00 262.8 F 23.3
32:00 272.4 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.8
1:00 271.9 F 30.2
2:00 272.2 F 30.0
3:00 272.1 F 29.8
4:00 271.9 F 29.8

Min 271.7 F 29.7
Max 272.4 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 215.0 F 0.0
10:00 206.0 F 0.0
15:00 190.2 F 0.0
20:00 179.3 F 0.0
25:00 188.1 F 0.0
30:00 192.0 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 33:11
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:11:48
CYCLE COMPLETE

03/13/2025 10:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____