
Midmark M11 - v1.0.0
Total Cycles: 2572

Name: RITTER 1
Sterilizer ID: V2178198
11/16/2022 06:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.8 F 0.0
2:00 193.6 F 4.5
4:00 211.2 F 10.0
6:00 224.5 F 5.9
8:00 234.7 F 11.4
10:00 236.1 F 8.9
12:00 238.4 F 9.4
14:00 249.1 F 14.8
16:00 261.9 F 22.0
18:00 272.3 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.5 F 29.0
2:00 271.3 F 28.4
3:00 272.3 F 28.8
4:00 271.3 F 28.5

Min 271.2 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 237.1 F 0.0
10:00 227.4 F 0.0
15:00 228.3 F 0.0
20:00 226.5 F 0.0
25:00 212.5 F 0.0
30:00 213.5 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 18:05
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:32
CYCLE COMPLETE

11/16/2022 07:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____