

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3529

Name: RITTER2  
Sterilizer ID: V2313228  
04/15/2024 03:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 113.0 F 0.0  
2:00 117.8 F 0.7  
4:00 140.8 F 3.3  
6:00 167.2 F 6.6  
8:00 189.5 F 10.4  
10:00 206.6 F 6.4  
12:00 219.8 F 8.0  
14:00 230.9 F 12.5  
16:00 232.0 F 9.4  
18:00 238.9 F 10.0  
20:00 245.3 F 13.5  
22:00 256.4 F 19.6  
24:00 266.7 F 26.2  
26:00 272.9 F 30.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.9  
1:00 271.5 F 29.8  
2:00 272.2 F 30.1  
3:00 272.2 F 30.3  
4:00 271.7 F 30.4

Min 271.1 F 29.8  
Max 272.7 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.7 F 0.0  
5:00 231.1 F 0.0  
10:00 195.8 F 0.0  
15:00 185.3 F 0.0  
20:00 181.1 F 0.0  
25:00 180.1 F 0.0  
30:00 179.8 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 26:22  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:05:05  
CYCLE COMPLETE

04/15/2024 04:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_