
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001495
12:40 10/13/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:11
PRESSURIZING 7:04
STERILIZING 8:38
136.0 C 220kPa 8:38
136.5 C 223kPa 9:08
135.9 C 220kPa 9:38
136.4 C 221kPa 10:08
136.4 C 223kPa 10:38
136.5 C 223kPa 11:08
136.4 C 222kPa 11:38
136.4 C 223kPa 12:08
VENTING 12:08
AIR DRYING 13:14
CYCLE COMPLETE 43:14

Digital Signature #
749F7F586C8F072C
