

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4688

Name: RITTER 1  
Sterilizer ID: V2178198  
06/28/2024 10:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.1 F 0.0  
2:00 139.4 F 2.0  
4:00 150.5 F 4.8  
6:00 174.7 F 8.2  
8:00 202.3 F 11.6  
10:00 210.5 F 4.4  
12:00 222.1 F 8.2  
14:00 230.5 F 12.5  
16:00 232.4 F 8.6  
18:00 243.5 F 12.7  
20:00 242.4 F 11.7  
22:00 253.7 F 17.4  
24:00 263.7 F 23.4  
26:00 272.5 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.5 F 29.3  
2:00 272.2 F 28.9  
3:00 271.2 F 28.4  
4:00 272.0 F 29.2

Min 271.2 F 28.4  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.1 F 0.0  
5:00 213.2 F 0.0  
10:00 213.6 F 0.0  
15:00 218.8 F 0.0  
20:00 215.5 F 0.0  
25:00 215.2 F 0.0  
30:00 217.3 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 26:17  
STERILIZING: 4:00  
VENTING: 3:54  
DRYING: 30:00  
TOTAL CYCLE: 01:05:06  
CYCLE COMPLETE

06/28/2024 11:56 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_