
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 008025
11:28 05/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:17
PRESSURIZING 3:18
STERILIZING 4:18
136.0 C 220kPa 4:18
135.9 C 219kPa 4:48
136.3 C 223kPa 5:18
136.4 C 224kPa 5:48
136.4 C 224kPa 6:18
136.4 C 224kPa 6:48
136.4 C 223kPa 7:18
136.1 C 221kPa 7:48
136.1 C 221kPa 7:48
VENTING 7:48
AIR DRYING 8:42
CYCLE COMPLETE 16:25

Digital Signature #
136E9E63244CCF73
