
Midmark M11 - v1.0.0
Total Cycles: 4245

Name: RITTER 1
Sterilizer ID: V2178198
01/29/2024 04:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 215.2 F 0.0
2:00 212.2 F 5.9
4:00 222.0 F 13.4
6:00 235.5 F 10.8
8:00 238.7 F 10.0
10:00 239.1 F 9.8
12:00 256.5 F 18.5
14:00 259.4 F 20.2
16:00 272.6 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 272.1 F 28.1
2:00 272.8 F 28.5
3:00 272.9 F 28.8
4:00 271.7 F 27.9

Min 271.6 F 27.8
Max 273.0 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 220.5 F 0.0
5:00 220.0 F 0.0
Dry Cycle Aborted

FILLING: 2:37
HEATING: 16:13
STERILIZING: 4:00
VENTING: 3:10
DRYING: 6:02
TOTAL CYCLE: 00:32:02

01/29/2024 05:16 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____