
Midmark M11 - v1.0.0
Total Cycles: 3585

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2023 10:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.2 F 0.0
2:00 187.1 F 2.1
4:00 202.0 F 5.6
6:00 216.4 F 10.6
8:00 225.6 F 5.7
10:00 233.3 F 9.8
12:00 238.8 F 9.2
14:00 242.4 F 12.2
16:00 240.8 F 10.6
18:00 253.4 F 16.9
20:00 264.5 F 23.4
22:00 271.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.7 F 28.8
2:00 271.7 F 28.1
3:00 272.0 F 28.7
4:00 272.4 F 28.6

Min 271.3 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 215.5 F 0.0
10:00 228.1 F 0.0
15:00 228.0 F 0.0
20:00 228.4 F 0.0
25:00 230.7 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 22:01
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:40
CYCLE COMPLETE

07/26/2023 11:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____