

-----  
Midmark M11 - v1.0.5  
Total Cycles: 431

Name: RITTER2  
Sterilizer ID: V2313228  
10/08/2021 07:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.4 F 0.0  
2:00 175.2 F 2.0  
4:00 181.2 F 6.3  
6:00 203.0 F 11.8  
8:00 221.4 F 7.3  
10:00 234.8 F 13.0  
12:00 238.4 F 10.7  
14:00 237.1 F 9.4  
16:00 250.9 F 16.1  
18:00 262.5 F 22.8  
20:00 272.6 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.9  
1:00 272.1 F 29.1  
2:00 272.3 F 29.5  
3:00 271.5 F 29.3  
4:00 271.8 F 28.9

Min 271.2 F 28.7  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.7 F 0.0  
5:00 228.0 F 0.0  
10:00 218.9 F 0.0  
15:00 220.4 F 0.0  
20:00 215.2 F 0.0  
25:00 218.8 F 0.0  
30:00 209.0 F 0.0  
Drying Complete

FILLING: 2:10  
HEATING: 20:16  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:59:32  
CYCLE COMPLETE

10/08/2021 08:55 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02