
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004073
9:26 21/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:47
PRESSURIZING 1:52
STERILIZING 3:01
136.4 C 223kPa 3:01
136.4 C 224kPa 3:31
136.2 C 221kPa 4:01
136.2 C 222kPa 4:31
136.4 C 225kPa 5:01
136.5 C 223kPa 5:31
136.4 C 224kPa 6:01
136.4 C 223kPa 6:31
136.4 C 223kPa 6:31
VENTING 6:31
AIR DRYING 7:18
CYCLE COMPLETE 7:20

Digital Signature #
DA87D815FDEA8907
