
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000716
9:00 21/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:18
STERILIZING 4:25
136.0 C 227kPa 4:25
136.2 C 230kPa 4:55
136.4 C 231kPa 5:25
136.4 C 233kPa 5:55
136.4 C 233kPa 6:25
136.4 C 232kPa 6:55
136.4 C 232kPa 7:25
136.4 C 233kPa 7:55
136.4 C 233kPa 7:55
VENTING 7:55
AIR DRYING 8:50
CYCLE COMPLETE 8:51

Digital Signature #
CD67E14FE1A20893
